



## Short audit physical aerosol properties – Checklist

Version 2.0 AW/TT 2014

Station name: \_\_\_\_\_

Date: \_\_\_\_\_

Auditors: \_\_\_\_\_

- Manuals for instruments available on site    yes                    no
- Written logbooks for each instrument        yes                    no
- General impression                    excellent            good            fair            poor

Comments: \_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Note: All flow rates of inlet and instruments should be measured!

Type of reference flow meter: \_\_\_\_\_                    Serial number: \_\_\_\_\_



World Calibration Centre  
for Aerosol Physics

**Aerosol inlet:**

Inlet: PM10 NOAA other\_\_\_\_\_

Material: stainless steel conductive tubing other \_\_\_\_\_

Design: vertical \_\_\_\_\_meters horizontal \_\_\_\_\_meters bends \_\_\_\_\_

tube diameter: \_\_\_\_\_ mm flow rate: \_\_\_\_\_l/min

Calculated Reynolds number (use *aerocalc*): \_\_\_\_\_

Calculated residence time in tube: \_\_\_\_\_

Aerosol dryer: yes no type:\_\_\_\_\_

Relative humidity of aerosol: \_\_\_\_\_% RH not available

Inlet according to recommendations: yes no

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_



**Absorption:**

Aethalometer    MAAP    PSAP    Other: \_\_\_\_\_    n/a

Type: \_\_\_\_\_    Serial number: \_\_\_\_\_

Wavelengths: \_\_\_\_\_

Firmware version: \_\_\_\_\_

Software version: \_\_\_\_\_

Data format: \_\_\_\_\_ (e.g. scientific=12 for MAAP)

Last calibrated: \_\_\_\_\_ by \_\_\_\_\_

Nominal flow rate: \_\_\_\_\_ measured flow rate: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Flow rate indicated on front panel: \_\_\_\_\_

Indicated concentration with absolute filter: \_\_\_\_\_

Current transmission: \_\_\_\_\_ %

Last filter change: \_\_\_\_\_ at \_\_\_\_\_ % transmission

Condition of instrument:            excellent    good    fair    poor

Data submitted to data centre:    yes    no

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_



**Scattering Coefficients:**

Nephelometer: TSI Ecotech Radiance Research Other: \_\_\_\_\_ n/a

Type: \_\_\_\_\_ Serial number: \_\_\_\_\_

Wavelengths: \_\_\_\_\_

Firmware version: \_\_\_\_\_

Software version: \_\_\_\_\_

Gases for span check: CO<sub>2</sub> SF<sub>6</sub> n/a

Last zero check: \_\_\_\_\_ last span check: \_\_\_\_\_

Last calibrated: \_\_\_\_\_ by \_\_\_\_\_

If possible ask for calibration of the instrument by station personnel.

Compare new calibration constants with previous calibrations recorded in logbooks.

Nominal flow rate: \_\_\_\_\_ Measured flow rate: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Condition of instrument: excellent good fair poor

Data submitted to data centre: yes no

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_







World Calibration Centre  
for Aerosol Physics

**Particle number size distribution:**

Type of instrument: \_\_\_\_\_ Manufacturer: \_\_\_\_\_ S/N: \_\_\_\_\_

Type of charger \_\_\_\_\_ Nominal Activity \_\_\_\_\_ Date manufactured \_\_\_\_\_

Aerosol dryer:    yes    no    Type \_\_\_\_\_

Sheath air dryer:    yes    no    Type \_\_\_\_\_

Temperature sensor aerosol:            yes            no

Temperature sensor sheath air:        yes            no

Humidity sensor aerosol:                yes            no

Humidity sensor sheath air:            yes            no

Nominal sheath air flow rate \_\_\_\_\_ l/min

Measured sheath air flow rate: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Nominal aerosol flow rate \_\_\_\_\_ l/min

Measured aerosol flow rate: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Indicated concentration with absolute filter: \_\_\_\_\_

Instrument built according to recommendations:    yes    no

Condition of instrument:                    excellent    good    fair    poor

Data submitted to data centre:            yes            no

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_



World Calibration Centre  
for Aerosol Physics

**Particle number concentration**

Type of instrument: \_\_\_\_\_ Manufacturer: \_\_\_\_\_ S/N: \_\_\_\_\_

Nominal \_\_\_\_\_ flow rate: \_\_\_\_\_ measured: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Last flow calibration: \_\_\_\_\_

Last efficiency calibration: \_\_\_\_\_ Material : \_\_\_\_\_ Which lab: \_\_\_\_\_

Indicated concentration with absolute filter: \_\_\_\_\_

Condition of instrument:            excellent   good   fair   poor

Nominal \_\_\_\_\_ flow rate: \_\_\_\_\_ measured: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Last flow calibration: \_\_\_\_\_

Last efficiency calibration: \_\_\_\_\_ Material : \_\_\_\_\_ Which lab: \_\_\_\_\_

Indicated concentration with absolute filter: \_\_\_\_\_

Condition of instrument:            excellent   good   fair   poor

Nominal \_\_\_\_\_ flow rate: \_\_\_\_\_ measured: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Last flow calibration: \_\_\_\_\_

Last efficiency calibration: \_\_\_\_\_ Material : \_\_\_\_\_ Which lab: \_\_\_\_\_

Indicated concentration with absolute filter: \_\_\_\_\_

Condition of instrument:            excellent   good   fair   poor

Data submitted to data centre:        yes        no

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_





World Calibration Centre  
for Aerosol Physics

**Cloud condensation nuclei concentration**

Type of instrument: \_\_\_\_\_ Manufacturer: \_\_\_\_\_ S/N: \_\_\_\_\_

Supersaturations used: 0.1 0.2 0.3 0.5 1.0 other: \_\_\_\_\_

Software status lights green: yes no: \_\_\_\_\_

Aerosol \_\_\_\_\_ flow rate: \_\_\_\_\_ measured: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Total \_\_\_\_\_ flow rate: \_\_\_\_\_ measured: \_\_\_\_\_ at \_\_\_\_\_ hPa, \_\_\_\_\_ C

Indicated concentration with absolute filter: \_\_\_\_\_

Last flow calibration: \_\_\_\_\_

Last supersaturation calibration: \_\_\_\_\_ Material (NH<sub>4</sub>)<sub>2</sub>SO<sub>4</sub> other: \_\_\_\_\_

Last OPC calibration (optional): \_\_\_\_\_

Condition of instrument:                    excellent   good   fair   poor

Data submitted to data centre:            yes            no

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_



World Calibration Centre  
for Aerosol Physics

**General remarks:**

---

---

---

---

---

---

Date: \_\_\_\_\_

Signature: \_\_\_\_\_